Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co., Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Device Group:	Z121201 - Ophthalmology Assessment and Diagnosis Instruments
Medical Device:	Trial Lenses Set
Specification:	MT260, PL232, MT90, MT68
Manufacturer SRN:	CN-MF-000012433
Device Basic UDI-DI:	69450339TLFM
UMDNS Code:	17555
Risk class of the device:	Annex VIII, Class I with measuring function, rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex XI Part A
Notified Body:	TÜV SÜD Product Service GmbH Ridlerstraße 65, 80339 Munich, GERMANY
Identification number:	CE 0123
(EC)Certificate(s):	G21 069181 0010 Rev. 00
Exp. date:	26 Apr. 2028
Authorized Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Signature:	Mr. Ge Xiaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

LIANYUNGANG TIANNUO OPTICAL INSTRUMENT CO., LTD. 连云港天诺光学仪器有限公司

We, the manufacturer, hereby declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices. All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.

Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co.,Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Medical Device:	Prism Set
Specification:	Prism Bar (PHB45-PVB30)
Single Registration Number:	CN-MF-000012433
Device Basic UDI-DI:	69450339PBEM
UMDNS Code:	15875
Risk class of the device:	Class I, rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex II+III
Identification number:	CE
European Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Signature:	Mr. Ge Xiaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

We, the manufacturer, hereby declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices; All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.

Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co., Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Medical Device:	Cross Cylinder Set
Specification:	CC0.12, CC0.25, CC0.50, CC1.00
Manufacturer SRN:	CN-MF-000012433
Device Basic UDI-DI:	69450339TLFM
UMDNS Code:	17555
Risk class of the device:	Annex VIII, Class I, rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex II+III
Identification number:	CE
Authorized Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Signature:	Mr. Ge Xiaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

LIANYUNGANG TIANNUO OPTICAL INSTRUMENT CO., LTD. 连云港天诺光学仪器有限公司

We, the manufacturer, hereby declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices. All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.

Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co., Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Medical Device:	Exophthalmometer
Specification:	EM-IV
Manufacturer SRN:	CN-MF-000012433
Device Basic UDI-DI:	69450339EXEY
UMDNS Code:	16345
Risk class of the device:	Annex VIII, Class I , rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex II+III
Notified Body:	TÜV SÜD Product Service GmbH Ridlerstraße 65, 80339 Munich, GERMANY
Identification number:	CE
(EC)Certificate(s):	Q5 069181 0011 Rev. 00
Exp. date:	28 Feb. 2027
Authorized Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Signature:	Mr. Ge Xiaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

LIANYUNGANG TIANNUO OPTICAL INSTRUMENT CO., LTD. 连云港天诺光学仪器有限公司

We, the manufacturer, here with declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices. All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.

Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co., Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Medical Device:	Trial clip
Specification:	TC1
Manufacturer SRN:	CN-MF-000012433
Device Basic UDI-DI:	69450339TFF9
UMDNS Code:	15582
Risk class of the device:	Annex VIII, Class I with measuring function, rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex XI Part A
Notified Body:	TÜV SÜD Product Service GmbH Ridlerstraße 65, 80339 Munich, GERMANY
Identification number:	CE 0123
(EC)Certificate(s):	G21 069181 0010 Rev. 00
Exp. date:	26 Apr. 2028
Authorized Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Start of CE-marking:	18 Mar. 2019
Signature:	Mr. Gao Xiaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

LIANYUNGANG TIANNUO OPTICAL INSTRUMENT CO., LTD. 连云港天诺光学仪器有限公司

We, the manufacturer, here with declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices. All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.

Declaration of Conformity

Manufacturer:	Lianyungang TianNuo Optical Instrument Co., Ltd No.3 JinQiao Road, Dapu Industrial park 222000 Lianyungang, Jiangsu PEOPLE'S REPUBLIC OF CHINA
Device Group:	Z121201 - Ophthalmology Assessment and Diagnosis Instruments
Medical Device:	Trial Frame
Specification:	UTF-5080, UTF-5080Pro, PD48, PD56
Manufacturer SRN:	CN-MF-000012433
Device Basic UDI-DI:	69450339TFF9
UMDNS Code:	16466
Risk class of the device:	Annex VIII, Class I with measuring function, rule 1
Conformity assessment Route:	Regulation (EU) 2017/745 Annex XI Part A
Notified Body:	TÜV SÜD Product Service GmbH Ridlerstraße 65, 80339 Munich, GERMANY
Identification number:	CE 0123
(EC)Certificate(s):	G21 069181 0010 Rev. 00
Exp. date:	26 Apr. 2028
Authorized Representative:	Shanghai International Holding Corp. GmbH (Europe) Eiffestraße 80, 20537 Hamburg, GERMANY
Signature:	Mr. Gao Kaojun
Position:	General manager
Place:	Lianyungang, Jiangsu, China
Date of Declaration:	July 7th, 2026

LIANYUNGANG TIANNUO OPTICAL INSTRUMENT CO., LTD.

We, the manufacturer, hereby declare that the stated medical device meet the transposition into national law, the provisions of Regulation (EU) 2017/745 on Medical Devices. All supporting documentation is retained at the premises of the manufacturer, and the manufacturer is exclusively responsible for the Declaration of Conformity.